



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

| APPLICATION NO. | FILING DATE | FIRST NAMED INVENTOR | ATTORNEY DOCKET NO. | CONFIRMATION NO. |
|-----------------|-------------|----------------------|---------------------|------------------|
|-----------------|-------------|----------------------|---------------------|------------------|

10/562,648

05/02/2006

Mark A. Rydell

59673.1.12.8.1.1

8434

22859

7590

10/14/2009

INTELLECTUAL PROPERTY GROUP

FREDRIKSON & BYRON, P.A.

200 SOUTH SIXTH STREET, SUITE 4000

MINNEAPOLIS, MN 55402

EXAMINER

STEWART, ALVIN J

ART UNIT

PAPER NUMBER

3774

MAIL DATE

DELIVERY MODE

10/14/2009

PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/562,648

Applicant(s)

RYDELL ET AL.

Examiner

Alvin J. Stewart

Art Unit

3774

Period for Reply -- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 15 June 2009.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-7, 19, 20, 22, 24 and 25 is/are pending in the application.
- 4a) Of the above claim(s) _ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-7, 19, 20, 22, 24 and 25 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☒ The drawing(s) filed on 27 December 2005 is/are: a) ☒ accepted or b) ☐ objected to by the Examiner.
- Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
- Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☐ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date _____
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: _____

DETAILED ACTION

Response to Arguments

Applicant's arguments filed 06/15/09 have been fully considered but they are not persuasive.

For example, the Applicant's representative added new limitations on claim 1 disclosing that the "tibia being allowed to articulate across the first major surface". The Examiner wants to remind the Applicant's representative that those limitations are only functional language and the new entered paragraph does not add any structure limitations to the claim. The Examiner has to only find a reference capable of having a tibia being allowed to articulate across the first mayor surface and the Examiner believes that the above reference reads on the above claimed functional language.

Regarding claim 22, the Examiner believes that the reference clearly discloses a first mayor surface and a second mayor surface. With respect to the phrase: "substantially free of anchoring portions", the examiner believes that the prior art still reads on the claimed subject matter because the anchoring system of the Parks et al reference is substantially smaller than the anchoring system of other ankle joints, such as anchoring systems that comprise a stem that goes within the intramedullary canal of the femur. Elements 46 and 54(a) are very small anchoring system compare to others such as Patents 7,485,147; 7,314,488, etc...

The rest of the limitations are functional language.

For the reasons above, the Examiner believes that the rejection is proper. In order to overcome the rejection the Applicant's representative most add more structure limitations that can distinguished the prior art from the patent application.

Regarding claim 25, see element 58 clearly disclosing a bead shaped structure.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-4, 19, 20, 22, and 24-25 are rejected under 35 U.S.C. 102(e) as being anticipated by Parks et al US Patent 7,025,790 B2.

Parks et al discloses an implant (50) having a convex outer surface and a concave inner surface. Wherein the inner surface comprises external structures such as integral bead (see col. 9, lines 61-63 and col. 10, lines 14-18).

Regarding claim 23, element 50 does not disclose anchoring portions, such as pin, stems, etc...

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 5-7 are rejected under 35 U.S.C. 103(a) as being unpatentable over Parks et al US Patent 7,025,790 B2 in view of Schon US Patent 6,572,620.

Parks et al discloses the invention substantially as claimed. However, Parks et al does not an implant made of polyurethane and a talus-calcaneus implant having a posterior and anterior lips.

Schon disclose a plate (20) capable of being used as a talus-calcaneus implant having an anterior and posterior lips.

Regarding claims 8 & 9, it would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the Parks et al reference and add another implant as the plate (20) shown in the Schon reference in order to replace another injured part of the patient's body.

Regarding claims 5-7, It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the polyethylene material of the Parks et al reference with the polyurethane material, since it has been held to be within the general skill of a worker in the art to select a known material on the basis of its suitability for the intended use as a matter of obvious design choice. *In re Leshin*, 125 USPQ 416.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Alvin J. Stewart whose telephone number is 571-272-4760. The examiner can normally be reached on Monday-Friday 7:00AM-5:30PM(1 Friday B-week off).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Dave Isabella can be reached on 571-272-4749. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Alvin J Stewart/
Primary Examiner, Art Unit 3774

October 12, 2009.